


NOV 08 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130404 ✓
Name	John W. Carroll	ID#	6775 ✓
		Date	10/30/2010 ✓
A	Agency	Fort Greely Police Dept. BIG4	Phone # 907-873-9120
	Instrument Location	Bldg. 663 booking room Fort Greely, Alaska 99731	
B	Alco S/N	X172958 ✓	Target Value .087 ✓
		High Pressure	790
	Alco Test Values	.084 ✓	.084 ✓
		1 st Alco	2 nd Alco
	Signature		
(OVER)			


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of DEC, 2010.

 (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 06/03



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130404

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

OCTOBER 30, 2010

OPERATOR'S NAME:

CARROLL, JOHN W

OPERATOR'S NUMBER: 6775

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 00000000

DEPT/AGENCY: FORT GREELY PD BIG4

CASE/REPORT: 00-0000

TEST TYPE: V

ALCO TARGET VALUE: .087

ALCO S/N: X172958

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 28.03 in

ALCO TARGET .081 21:34

BLANK TEST .000 21:35

INTERNAL STANDARD VERIFIED 21:35

ALCO TV 28.03 in .084 21:35

BLANK TEST .000 21:36

SUBJECT SAMPLE .000 21:37

BLANK TEST .000 21:38

ALCO TV 28.03 in .084 21:38

BLANK TEST .000 21:39

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130404

OCTOBER 30, 2010

TIME 21:30

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 28.03 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!""%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~